



Guam Memorial Hospital Authority Aturidåt Espetåt Mimuriåt Guåhan



850 GOV. CARLOS CAMACHO ROAD
OKA, TAMUNING, GUAM 96911
TEL: 647-2444 or 647-2330
FAX: (671) 649-0145

June 07, 2018

AMENDMENT #2 FOR GMHA IFB 003-2018

Infant Warmer

TO: All prospective bidders,

This Amendment is being issued to clarify the following:

1) Bidder-(Medpharm): Upon review of the Technical Specifications, it is apparent they are based on GE Panda Plus. The hands-free alarm silencing feature is a functionality that offers no clinical significance and is a restrictive GE feature.

GMHA:


Hands free alarm is of clinical significance and is not restrictive to GE. See Okuman and Novos brand warmers who also feature the hands free alarm.

2) Bidder-(Medpharm): 2. Biomedical Technician/Clinical Training at the company training facility - It will be too expensive for 2 Units of Infant Warmers. Can we provide local/on-site training?

GMHA:

If the provider has a local company training facility then that would suffice. The training should be conducted by the equipment certified trainer and can certify attendees as certified technicians.

If any other questions/clarifications please send via telefax addressed to Mr. Peterjohn D. Camacho, MPH


Dolores E. Pangelinan,
Hospital Supply Management Admin.

Acknowledgment of Receipt: Return acknowledgment to fax number 649-3640.

Company

Print Name

Signature

Date