



Guam Memorial Hospital Authority Aturidåt Espetåt Mimuriåt Guåhan



850 GOV. CARLOS CAMACHO ROAD
OKA, TAMUNING, GUAM 96911
TEL: 647-2444 or 647-2330
FAX: (671) 649-0145

June 12, 2018

AMENDMENT #3 FOR GMHA IFB 003-2018

Infant Warmer

TO: All prospective bidders,

This Amendment is being issued to clarify the following:

1) Bidder- (Medpharm): Pulse Oximetry-Please specify if Masimo or Nellcor?

GMHA:

L&D currently uses Nonin for our portable O2 monitors.

2) Bidder-(Medpharm): Will the Warmers stay in the room or be moved around from one room to another—this would require tank mounts.

GMHA:


We will move the warmer around so tank mounts will be needed.

3) Bidder-(Medpharm): What type of wall gas connections do you have?

GMHA:

We have oxygen and air connections to use with the blender.

If any other questions/clarifications please send via telefax addressed to Mr. Peterjohn D. Camacho, MPH


Dolores F. Pangelinan,
Hospital Supply Management Admin.

Acknowledgment of Receipt: Return acknowledgment to fax number 649-3640.

Company

Print Name

Signature

Date