

**GUAM MEMORIAL HOSPITAL AUTHORITY
850 GOV. CARLOS G. CAMACHO ROAD
TAMUNING, GUAM 96913
PHONE: 671-647-2165
EMAIL: materials.mgmt@gmha.org**

TRANSMITTAL SHEET

TO: Prospective Bidders
FROM: Hospital Materials Management Administrator
DATE: September 13, 2024
SUBJECT: **IFB-020-2024- Purchase of Fluoroscopy X-Ray System for the GMHA Radiology Department. Amendment No. 4**
PAGES: 4 including cover sheet

NOTES:

An acknowledgement via a return email would be appreciated as soon as possible.

DATE / VENDOR ACKNOWLEDGEMENT

FACSIMILE CONFIDENTIALITY NOTICE

These documents accompanying this transmission contain confidential information that is legally privileged. This information is intended only for the use of the individual or entity named above. The authorized recipient of this information is prohibited from disclosing this information to any other party unless required to do so by law or regulation and is required to destroy the information after its stated need has been fulfilled.

If you are not the intended recipient, you are hereby notified that any disclosure, copying, distribution, or action taken in reliance on the contents of these documents is strictly prohibited. If you have received this information in error, please notify the sender immediately and arrange for the return or destruction of these documents.



GUAM MEMORIAL HOSPITAL AUTHORITY

ATURIDÁT ESPETÁT MIMURIÁT GUÁHÁN

850 Governor Carlos Camacho Road, Tamuning, Guam 96913
Operator: (671) 647-2330 or 2552 | Fax: (671) 649-5508



AMENDMENT #4 GMHA IFB 020-2024 Purchase of Fluoroscopy X-Ray System

This amendment is in response to clarification from JMI.

Clarification: Does GMHA require a wall bucky?

Response: No.

Clarification: Is the requirement on detectors for a total of 2 or 3 detectors? The detector in the Fluoro table is fixed which cannot be removed.

Response: No.

Clarification: There is no space to put a very large battery back up system to maintain the full x ray functionality in the requirement. Hence, the back up power or 6 KVA UPS can only power up the Workstation when there is a power outage. Please confirm if GMHA still wants a full UPS to power the full Fluoro system or not because there is no space for the large UPS cabinet.

Response: No, UPS for workstation is fine.

Clarification: Please define technical assessment in the price breakdown. I'm sorry but I don't understand why there is a line item to put a price on it.

Response: Attached to this amendment is the corrected Price/Proposal Breakdown that removes the "Technical Assessment" that was placed on breakdown in error.

Clarification: Please advise if all the tiles for the whole examination room where the Fluoro table is located is required to be replaced or not.

Response: Yes, flooring needs to be replaced.

Please acknowledge receipt of this amendment by signing and sending back to Materials Management by Fax at 649-3640 or email to materials.mgmt@gmha.org.

If you have any questions, please feel free to address your letter to Lillian Perez-Posadas MN, RN, and fax to the Materials Management Office at 649-3640 or email to the materials.mgmt@gmha.org.

Sincerely;

DOLORES PANGELINAN

Hospital Materials Management Administrator

ACKNOWLEDGMENT:

PRINT NAME

SIGNATURE

DATE

PRICE/PROPOSAL BREAKDOWN

Fluoroscopy X-ray System, including removal and disposal of old equipment, installation of new equipment, training and installation of related and associated system, system operational check, testing and commissioning, and 2-year warranty.

\$ _____

Area/Room Upgrade

\$ _____

Others/Miscellaneous (Please Specify)

\$ _____

TOTAL BID AMOUNT:

\$ _____

15% OF BID AMOUNT:

\$ _____

NOTE: All items are for Guam Memorial Hospital Authority (GMHA). All prices must be inclusive of shipping cost and/or installation cost.

These specifications have been prepared by: Leland Guiao, Chief of Radiology and Adrian Manuel, Acting Facilities Maintenance Manager.

APPROVED BY:

LILLIAN PEREZ-POSADAS, MN, RN
Hospital Administrator/CEO/Chief Procurement Officer

Date